

covidien kendall scd 700 series user manual

Covidien Kendall SCD 700 Series User Manual

The Covidien Kendall SCD 700 Series represents a significant advancement in the field of compression therapy, designed to prevent deep vein thrombosis (DVT) and enhance patient recovery. This user manual serves as a comprehensive guide for healthcare professionals and caregivers on how to effectively operate, maintain, and troubleshoot the device. By adhering to the guidelines outlined in this manual, users can ensure optimal performance and patient safety.

Introduction to the SCD 700 Series

The SCD 700 Series is a pneumatic compression device that utilizes sequential compression to promote venous blood flow and prevent the formation of blood clots. The system is equipped with various cuffs that can be applied to different parts of the body, making it versatile for a range of patients, including those undergoing surgery, immobilized patients, and those with limited mobility.

Key Features

- Sequential Compression: The device inflates and deflates cuffs in a specific sequence to mimic natural leg muscle movement.
- Adjustable Pressure Settings: Users can modify pressure levels to accommodate individual patient needs.
- User-Friendly Interface: The SCD 700 Series features an intuitive control panel for easy operation.
- Portable Design: Lightweight and compact, making it suitable for hospital and home use.
- Safety Alarms: Equipped with audible alerts for device malfunctions and improper cuff application.

Components of the SCD 700 Series

Understanding the various components of the SCD 700 Series is crucial for effective usage. Below is a breakdown of the main parts:

1. Control Unit: The heart of the system, where settings are adjusted and monitored.

2. Cuffs: These are available in various sizes and are designed to fit snugly around the patient's legs or arms.
3. Tubing: Connects the cuffs to the control unit for air inflation and deflation.
4. Power Supply: Provides necessary electrical power to the device.
5. User Manual: A vital resource for understanding device operation and maintenance.

Operating Instructions

To ensure safe and effective use of the SCD 700 Series, follow these operating instructions:

Preparation

1. Inspect the Device: Before use, check for any visible damage or wear on the device and accessories.
2. Select the Appropriate Cuff Size: Measure the patient's limb to ensure a proper fit. Cuffs that are too tight or too loose can affect performance.
3. Connect the Cuffs: Attach the cuffs to the control unit using the provided tubing.

Setup Procedure

1. Turn on the Control Unit: Plug the unit into a power outlet and press the power button.
2. Adjust Pressure Settings: Use the control panel to set the desired pressure level based on the patient's comfort and physician's recommendations.
3. Position the Cuffs: Wrap the cuffs around the patient's legs or arms, ensuring they are snug but not restrictive.
4. Initiate Compression Therapy: Start the therapy by pressing the start button on the control unit.

Monitoring and Adjustments

During the operation of the SCD 700 Series, continuous monitoring is essential:

- Observe Patient Comfort: Regularly check on the patient to ensure they are comfortable and that the cuffs are not causing any discomfort.
- Monitor Device Performance: Keep an eye on the control unit for any alarms or alerts indicating issues.

- **Adjust Settings as Needed:** If the patient experiences discomfort, adjust the pressure settings accordingly.

Maintenance and Care

Proper maintenance is vital to ensure the longevity and effectiveness of the SCD 700 Series:

Daily Maintenance

1. **Clean the Cuffs:** After each use, clean the cuffs with a damp cloth and mild detergent.
2. **Inspect Tubing:** Check for any kinks or blockages in the tubing.

Periodic Maintenance

1. **Calibrate the Control Unit:** Follow manufacturer guidelines to ensure the device remains accurate.
2. **Check for Software Updates:** Regularly update the control unit software to enhance functionality and performance.

Storage Guidelines

- **Store in a Cool, Dry Place:** Avoid direct sunlight and excessive humidity.
- **Keep Cuffs Flat:** Ensure that cuffs are stored flat to prevent creasing and damage.

Troubleshooting Common Issues

Despite its robust design, users may encounter some common issues while using the SCD 700 Series. Here are typical problems and their solutions:

Device Does Not Power On

- Ensure the power supply is correctly plugged in.
- Check the power outlet for functionality.

Alarms Sounding

- Confirm that the cuffs are properly connected and positioned.
- Check for any blockages in the tubing.

Insufficient Compression

- Verify that the pressure settings are correctly adjusted.
- Inspect the cuffs for any damage or wear.

Conclusion

The Covidien Kendall SCD 700 Series is an essential tool in modern healthcare, providing effective compression therapy for patients at risk of DVT. By following the guidelines in this user manual, healthcare providers can ensure the proper use, maintenance, and troubleshooting of the device. This not only enhances patient care but also promotes quicker recovery times. Always remember to consult with medical professionals for personalized advice concerning the use of the SCD 700 Series, and stay updated with the latest practices and recommendations in compression therapy.

Frequently Asked Questions

What is the primary function of the Covidien Kendall SCD 700 Series?

The primary function of the Covidien Kendall SCD 700 Series is to provide intermittent pneumatic compression to help prevent deep vein thrombosis (DVT) in patients.

Where can I find the user manual for the Covidien Kendall SCD 700 Series?

The user manual for the Covidien Kendall SCD 700 Series can typically be found on the official Medtronic (Covidien) website under the product support section or by contacting their customer service.

What are the key features of the Covidien Kendall SCD 700 Series?

Key features of the Covidien Kendall SCD 700 Series include adjustable pressure settings, multiple treatment modes, and a user-friendly interface

for ease of use.

How do I set up the Covidien Kendall SCD 700 Series for use?

To set up the Covidien Kendall SCD 700 Series, connect the compression sleeve to the device, plug it into a power source, and select the desired settings according to the user manual.

What should I do if the Covidien Kendall SCD 700 Series displays an error message?

If the Covidien Kendall SCD 700 Series displays an error message, refer to the troubleshooting section of the user manual for specific guidance on how to resolve the issue.

Is the Covidien Kendall SCD 700 Series suitable for all patients?

The Covidien Kendall SCD 700 Series is not suitable for all patients; it is important to consult with a healthcare professional to determine if it is appropriate based on individual patient conditions.

Can the Covidien Kendall SCD 700 Series be used at home?

Yes, the Covidien Kendall SCD 700 Series can be used at home, but it should be done under the guidance of a healthcare provider to ensure proper use and monitoring.

What are the maintenance requirements for the Covidien Kendall SCD 700 Series?

Maintenance requirements for the Covidien Kendall SCD 700 Series include regular cleaning of the device and sleeves, checking for wear and tear, and ensuring all connections are secure.

[Covidien Kendall Scd 700 Series User Manual](#)

Find other PDF articles:

<https://staging.liftfoils.com/archive-ga-23-17/pdf?ID=ars80-6558&title=developing-talent-in-young-people.pdf>

Covidien Kendall Scd 700 Series User Manual

Back to Home: <https://staging.liftfoils.com>